

F-NTP-2020
11/17/20



Republic of the Philippines
PROVINCE OF BOHOL
City of Tagbilaran



OFFICE OF THE GOVERNOR
NOTICE TO PROCEED

November 17, 2020

MS. STELLA MARIE C. ANCLA
Proprietor/Manager
DEXTEL TRADING
0410 Tamblot Street, Cogon District,
Tagbilaran City

Dear Ms. Ancla,

Please be informed that the attached **Purchase Order No. PHO – 2020 – 331** dated **November 17, 2020** has been approved. Notice is hereby given to immediately deliver the items enumerated herein within **NINETEEN (19) DAYS** upon receipt of this notice and delivered to Identified Recipient Hospitals.

In view thereof, you shall be responsible for performing the delivery of items & services under the terms and conditions of the Purchase Order and in accordance with the Implementation Schedule.

Please acknowledge receipt and acceptance of this notice and be guided accordingly.

Truly yours,

ATTY. ARTHUR C. YAP
Governor

ATTY. KATHYRIN FE D. PIOQUINTO
Provincial Administrator

Date Approved: NOV 25 2020

I acknowledge receipt of this Notice:

Date Released to Supplier: NOV 25 2020 : _____
Authorized Signature
(Name of the Representative of the Bidder)

Date Returned by Supplier: NOV 25 2020 : _____
Authorized Signature
(Name of the Representative of the Bidder)



PURCHASE ORDER

PROVINCIAL HEALTH OFFICE, NEW CAPITOL BUILDING, COGON DISTRICT, TAGBILARAN CITY
Agency/Procuring Entity

Supplier :	DEXTEL TRADING	D.O. No. :	P.O. # - PHO - 2020 - 331
Address :	0410 Tamblot Street, Cogon District, Tagbilaran City	Date :	November 17, 2020
Telephone No.:		Mode of Procurement :	PUBLIC BIDDING
TIN :			

Gentlemen/Mesdames:

Please furnish this Office the following articles subject to the terms and conditions contained herein:

Place of Delivery:	IDENTIFIED RECIPIENT HOSPITALS	Delivery Term: Please Read Terms & Conditions at the
Date of Delivery :	NINETEEN (19) DAYS UPON THE APPROVAL OF P.O./ NOTICE TO PROCEED	Payment Term: Upon Completion of Delivery
Warranty Period:	ONE (1) YEAR	

Item No.	QTY	UNIT	DESCRIPTION OF SUPPLIES	UNIT COST	AMOUNT
1	2	unit	Electrolyte Analyzer - G.E 300 · ISE (Ion Selectivity Electrode) · Low Reagent consumption · Maintenance free electrodes · Automatic reagent replacement · Simple and reliable design · Test: K+, Na+, Cl-, iCa2+, nCa2+, TCa, pH Throughput: 60 tests per hour Connectivity: RS232 Display: 320X240 touch screen Operating Environment: Temperature: 20%-80%, Humidity: 35%-85% Power requirement: 110-240AV+10%,50/60 Hz 80W Dimension: 600mm (H) x680mm (W) x550mm (D) Weight: 8kg	417,100.00	834,200.00
2	1	unit	X-Ray, digital, 500MA - BROWNER standard output power : 50kW 1.2 mA range : 10-630mA 1.3 Tube voltage range : 40-150kV 1.4 mAs range : 0.1-630mAs 1.5 Load time range : 0.001s~6.3s 1.6 Input power voltage: 380/400VAC, ±10%, 50/60hz, 3phase; 1.7 High voltage transformer structure : The power frequency 1.8 High voltage frequency: 25kHz ± 5% 2.X-ray tube 2.1 Focus range : 0.6/1.2mm 2.2 Anode type: rotating anode, rotation speed: 3200 rpm 2.3 The maximum voltage : 150kV 2.4 Anode angle 12° 2.5 Anode heat capacity: 150kHu 2.6 Tube heating capacity: 1250kHu 2.7 Anode input power: big focus 50kW, small focus 22kW 2.8 The inherent filtration : 1.3mm Al/75kV 3. Collimator 3.1 The inherent filtration : 1.2 mm Al /75 kV ; 3.2 Collimator is equipped with LED lights (Bright 30S are automatically extinguished) 3.3 Collimator is built into measuring ruler; 4.1 Tube Tilting Angle : ±180° 4.2 Tube Vertical Rotation Angle:-10°~90° 4.3 Post their rotary Angle: - 180 °~+ 180 °,each 90 ° automatic positioning; 4.4 Vertical stroke of column along bed: more than1700mm 4.5 The head of the machine is up and down the column: it is greater than 1200mm 4.6 The Collimator of the limiting device : 220 ~ 2000mm 4.7 Height of camera bed: 660mm + 5% 4.8 The vertical movement of the bed surface is greater than 900mm 4.9 The transverse movement of the bed surface is greater than 250mm 4.10 The chest frame height is 2100mm plus or minus 5% 4.11 The motion stroke of the chest frame detector is greater than 1450mm 4.12 The longitudinal movement of the box is greater than 520mm 4.13 Rotation Angle of the beam limiting device: plus or minus 360°; WIFI connection , Amorphous silicon,whole panel technology 5.2 Detector valid exposure size : 430x430mm 5.3 Pixel size : 139 µm 5.4 The effective camera size of the detector : 17x17inch 5.5 Gathering matrix : 3072x3072 pixels	5,820,000.00	5,820,000.00

			5.6 Spatial resolution:3.6lp/mm 5.7 Pixel gray level : :14bit 5.8 Data output: wired/wireless; CPU:3.2GHz 6.2 Memory capacity : 4GB 6.3 Hard drive capacity : 500 GB 6.4 Resolution: 1280 x 1024 6.5 Display size : The 21inch color LCD capacitive touch screen, multi-touch 6.6 64 bit Windows operation system;		
			Patient registration: acquisition, find the patient data, to prepare for the photography 7.2 Radio graphic examination : setting system parameters, exposure photography and get the image 7.3 Images display, image analysis, processing and output 7.4 Data management: managing patient data, leading in, leading out , batch file, etc in accordance with		
			DICOM3.0 7.5 Configure calibration: the system components such as flat-panel detector, such as tube for automatic configuration and calibration 7.6 System diagnosis: monitoring system status, diagnosis of the system fault; Frame size 8.1.1 Width : 1070 mm 8.1.2 Length : : 2200 mm 8.1.3 Height : 2140 mm 8.2 Lathe bed size 8.2.1 Width : 710 mm 8.2.2 Length : : 1250 mm 8.2.3 Height : 500 mm 8.3 Size of the bed 8.3.1 Width : 820 mm 8.3.2 Length : : 2120 mm 8.4 Total weight (no detector) : 700kg 8.5 Bed bearing: 200kg 8.6 Support automatic exposure dose control function; The machine warranty 1 year 9.2 1 set of complete operation and maintenance manual 9.3 Provide site application and service training 9.4 Free service hot line		
			resolution:3.6lp/mm 5.7 Pixel gray level : :14bit 5.8 Data output: wired/wireless; CPU:3.2GHz 6.2 Memory capacity : 4GB 6.3 Hard drive capacity :		
			500 GB 6.4 Resolution: 1280 x 1024 6.5 Display size : The 21inch color LCD capacitive touch screen, multi-touch 6.6 64 bit Windows operation system; Patient registration: acquisition, find the patient data, to prepare for the photography 7.2 Radio graphic examination : setting system		
			parameters, exposure photography and get the image 7.3 Images display, image analysis, processing and output 7.4 Data management: managing patient data, leading in, leading out , batch file, etc in accordance with DICOM3.0 7.5 Configure calibration: the system components		
			such as flat-panel detector, such as tube for automatic configuration and calibration 7.6 System diagnosis: monitoring system status, diagnosis of the system fault; Frame size 8.1.1 Width : 1070 mm 8.1.2 Length : : 2200 mm 8.1.3 Height : 2140 mm 8.2 Lathe bed size 8.2.1 Width : 710 mm 8.2.2 Length : : 1250 mm 8.2.3 Height : 500 mm 8.3 Size of the bed 8.3.1 Width : 820 mm 8.3.2 Length : : 2120 mm 8.4 Total weight (no detector) : 700kg 8.5 Bed bearing: 200kg 8.6 Support automatic exposure dose control function;		
			The machine warranty 1 year 9.2 1 set of complete operation and maintenance manual 9.3 Provide site application and service training 9.4 Free service hot line		
3	3	unit	ECG Machine, 12 Leads with analyzer - NEWTECH OR CHINA	77,600.00	232,800.00
			- 12 lead resting ECG with standard ECG measurements Backlit VGA LCD screen		
			- Full Alphanumeric keyboard		
			- High resolution thermal print head		
			-Automatic and Manual Operation		
			-Storage for up to 50 ECGs		
			- 1 year warranty		
4	1	unit	Dental Chair with Accessories	277,420.00	277,420.00
			Fully adjustable:		
			: Multifunction Foot control		
			*adjustable tilting backrest		
			*adjustable seat cushion height		
			Patient chair		
			: 24V DC electric motor		
			: High quality upholstery		
			: Ergonomically designed head and armrests		
			Dental Light:		
			page 2 of 7 pages . . .		
			Next page pls . . .		

			· LED lighting with toggle switch		
			DENTAL TRAY:		
			· 470X300 MM SPACE FOR DENTAL UNIT		
			· 3 CONTROL VALVES FOR AIR AND WATER		
			· 3 HIGH SPEED HANDPIECE AND CONNECTOR		
			· 3 WAY SYRINGE		
			· PERIPHERAL Xray viewer		
			· 12 FUNCTION CONTROL PANEL		
			· SALIVA EJECTOR		
			WATER UNIT		
			· Glass cuspidor bowl with 1 ceramic strainer		
			· Automatic flushing system		
			· Automatic cup filler with heater		
			· High Vacuum saliva ejector		
			· 95x150mm tray for small instruments		
			· 3-way syringe		
			· Water tank for distilled water for handpieces		
			· Water hose and hose tap connector for direct water line		
			CONTROL BOX		
			· MAIN CONTROL FOR AIR AND WATER		
			· Drainage for waste/saliva		
5	1	unit	Defibrillator with Monitor	485,000.00	485,000.00
			Display:		
			-Screen type: High resolution display		
			-Screen Type: 7 inches (17.7cm) diagonally		
			-Sweep Speed: 25mm/sec		
			-Information: Heart Rate, Lead/Pads, Alarm On/Off, SpO2, AED Functions and Prompts, Alarm Selection and Limits, Delivered Energy		
			Defibrillator		
			-Waveform: Biphasic		
			-Electrode Impedance Measurement Range: 0-0250 ohms ECG Monitoring		
			-Patient connection: 3-lead ECG Cable; 5-lead ECG cable, paddles		
			-Lead selection: Displayed on monitor, paddles; I, II, III, AVR, AVF, V		
			-ECG size: 0.25, 0.5, 1, 1.5, 2, 4 cm/m V display on monitor		
			-Heart Rate: 20-300 BPM		
			-Battery Type: Rechargeable; Ni-MH battery, 12V SpO2 Module		
			-Measurement Range: 30-100%, ±2% between 80%-90%, others ± 5%		
			-Alarm Time Accuracy: less than 12 sec		
			Recorder:		
			-Paper: 50mm thermal		
			-Speed: 12.5mm/sec, 25mm/sec, 50mm/sec, User selectable 6 seconds delay		
			-AED Mode		
			-AED Function: Auto analyze and charge x3 with programmable auto energy level selection, screen prompts and voice prompts. Showable Rhythms: Ventricular fibrillation, Ventricular, tachycardia; QRS complex wave duration		
			Charge Control: Control on device front panel, press key on paddle		
			Manual Mode		
			Energy Selection: Selectable at 2, 5, 7, 10, 20, 30, 50, 70, 100, 150, 200, 300, 360 joules, Synchronized Mode: Synchronizes defibrillator pulse to patient's R-wave, "SYNC" messages displayed on monitor		
6	1	unit	Fully Automated Hematology Analyzer, 5 Part - MINDRAY	2,910,000.00	2,910,000.00
			Advantages & Features		
			· 5-part diff, 23 parameters, 3 scatter grams, 3 histograms		
			· 40 samples /hour		
			· sample volume: 15µl		
			· two counting modes: whole blood & prediluted		
			· Supports whole blood mode for capillary sample		
			· 20,000 storage capacity		
			· Intuitive operation system via 10.4" touch screen		
			page 3 of 7 pages . . .		
			Next page pls . . .		

			· Laser scatter + Chemdye +Flow cytometry		
			· Independent channel & optical method for Basophil measurement		
			Parameters		
			WBC, Lym%, Mon%, Neu%, Bas%, Eos%, Lym#, Mon#, Neu#, Eos#, Bas#, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-CV, RDW-SD, PLT, MPV, PDW		
			3 histograms for WBC, RBC and PLT		
			3 scattergrams: 4-differential scattergram and 2 scattergrams for eosinophil		
			Sample Volume		
			Prediluted 20µL; Whole Blood 15µL; Capillary WB blood 15µL		
			Throughput		
			40 samples per hour		
			Carryover		
			WBC, RBC, HGB < 0.5%, PLT<1.0%		
			Connectivity		
			USB, LAN, COM		
			Supports Uni- or bi-directional LIS LAN Port supports HL7 protocol		
			Printout		
			External Laser printer / Inkjet printer, various printout formats and		
7	1	unit	Fully Automated Clinical Chemistry Analyzer - MINDRAY	2,080,650.00	2,080,650.00
			Advantages & Features:		
			· Grating optical system with 12 wavelengths		
			· High efficiency standalone mixing bar		
			· Throughput: 320 tests per hour for chemistry		
			· 8-step auto wash system		
			· Liquid level detection and tracking		
			· 150µl minimum reaction volume		
			· Highly compatible reagent system: reagents, QC & Calibrators ready for use		
			Sample Tests that can be performed:		
			ALT, AST, ALP, ACP, α-HBDH, LDH, CK, CK-MB, α-AMY, MSO, ChE, ADA, γ-GT MAO, Lip, Lap, NAG, TG, TC, HDL Chole, LDL Chole, Glu, Lactate, UREA, Cr, UA, TP, Alb, TBIL, DBIL, TBA, FMN, APO-AI, APOB, Lp(a), ASO, C3, C4, CRP, IgA, IgM, IgG, Microalbumin, RF, β 2MG, Myoglobin, TF, FIB, HbA1c, HbAO, U-Alb, BTR		
			Principles		
			Absorbance photometry, turbidimetry, Ion selective electrode technology		
			Methodology		
			End-point, fixed time, kinetic, optional ISE, single/dual reagent chemistries, monochromatic/ biochromatic, linear non-linear multipoint calibration		
			Throughput		
			Up to 320 tests/h, or 480 tests/h with ISE		
			Connectivity		
			RS232		
			Operating System		
			Windows® XP Professional/Home SP2 or above Windows® 7		
			Operating Environment		
			Temperature: 15°C~30°C		
			Humidity: 35%~85%		
			Power requirement		
			AC 200~240V, 50/60Hz, ≤1500VA or		
			AC 100~130V, 50/60Hz, ≤1500VA		
			Dimension		
			860mm (W) x700mm (D) x625mm (H)		
			Weight		
			130kg		
8	35	unit	Hospital Bed with mattress, Manual, double crank - STANDARD	33,950.00	1,188,250.00
			Powder Coated steel Head and foot board		
			Collapsed type aluminum alloy railing		
			Perforated powder coated steel bed platform		
			Noiseless 5" casters with brake		
9	1	unit	Delivery Table (Parturition) - STANDARD	388,000.00	388,000.00
			Size: 1800*500*800mm		
Page 4 of 7 pages ... Next page pls ...					

			Main technical parameters:		
			Back Section Raised from the Horizontal > 55°		
			Leg Section Lowered from the horizontal >85°		
10	1	unit	Autoclave, Heavy Duty, 300L	1,178,550.00	1,178,550.00
			Physical :Capacity : 300 lits 24"x40" volume power 220V. Control unit pressure switch vacuum pump diaphragm signle head; Features:Auto water fillin < auo exhaust< Liquid sterilization mode, Pre vacuum function< DRY function,vacuum release, dry time(Post vacuum 30mins redry time 10 mins.;		
			SAFETY DEVICES: Dual safety door lock* non operational mode of device when door is not properly closed. Overheat cut off switch. Water filling detection time*door unlock reminder*low water sensor with buzzer alarm*Breaer switch (non fuse type) *Emergency Exhaust switch *over heat/ pressure indicator		
11	3	unit	Doppler Apparatus	48,500.00	145,500.00
			Advanced real time auto correlationfetal heart rate processing		
			Active noise reduction for crystal clear heart sounds; Improved sensitivity for reliable detection in early gestation; Improved performance on difficult mums; LCD display of fetal heart rate including pulse indicator; Audio output for headphones; Built in loudspeaker; 3MHZ waterproof transducer suitable for use in a water birthing environment (FD3);		
			High sensitivity for optimum performance in antenatal screening; Advanced power management with auto switch off to conserve battery life; Long robust coiled cable enables optimum transducer position in all circumstances; Integral probe protection and storage; Easy clean probe for improved infection control		
12	5	unit	Emergency Cart, delivered fully assembles no tools required - STAINLESS	97,000.00	485,000.00
			Top constructed of solid high density polythethylene; 100* total replacement lifetime guarantee		
13	50	unit	Bedside Table - ABS STANDARD	4,656.00	232,800.00
			Advanced electrostatic plastic spray of surface , anti aging and anti-rust, Cabinet frame is welded by profile steel stable and reliable, ABD ENGINEERING PLASTIC, Length : 47.5 cm width: 47.5 cm heoght 75 cm		
14	2	unit	Portable X-Ray Machine - BROWINER	2,910,000.00	5,820,000.00
			standard output power : 5.6kW 1.2 mA range : 10-100mA 1.3 Tube voltage range : 40-125kV 1.4 mAs range : 0.1-200mAs 1.5 Load time range : 0.002s~4s 1.6 High voltage transformer structure : The power frequency 1.7 Focus range : 0.6/1.8mm 1.8 Anode types: fixed anode 1.9 The maximum voltage : 125kV 1.10 Anode angle 15° 1.11		
			The inherent filtration : 0.6mm Al/75kV 1.12 Anode heat capacity : 42kHu 2. Collimator 2.1 The inherent filtration : 1 mm Al /70 kV ; 2.2 Collimator is equipped with LED lights (Bright 30S are automatically extinguished) 2.3 Collimator is built into measuring ruler 3. The tube bracket structure 3.1 SID: 330 ~ 1650mm 3.2 Size after folding:1060*400*400mm 3.3		
			weight : 32KG 4. Detector (iRay) 4.1 WIFI connection, Amorphous Silicon oxide ,whole panel technology 4.2 The effective camera size of the detector : 350mm×430mm 4.3 Pixel size : 150 μm 4.4 Gathering matrix :		
			2304×2800pixels 4.5 Spatial resolution : 3.4lp/mm 4.6 Pixel gray level : 14bit 4.7 Detector cooling mode: natural cooling 5. Digital image workstation 5.1 CPU frequency : CORE i3 5.2 Memory capacity : 4GB 5.3 Hard drive capacity : 500 GB 5.4 64 bit Windows operation system 6. Image acquisition software 6.1 Patient registration: acquisition, find the patient data, to prepare for the photography 6.2 Images display, image analysis,		
			<i>page 5 of 7 pages ...</i>		
			<i>Next page pls ...</i>		

			processing and output 6.3 Data management: managing patient data, leading in, leading out, batch file, etc in accordance with DICOM3.0,5.4AH 6.4 System diagnosis: monitoring system status, diagnosis of the system fault 7. Battery 7.1 Power supply way: Lithium battery supply 7.2 The rated capacity:32.4V 7.3 Battery capacity: More than 100 exposures, (Exposure conditions: 50kV, 80mA, 20ms.)		
			7.4 Dual power supply, exposure under either wall current or lithium battery 7.5 Maximum charging 4 hours 7.6 The standby time is not less than 7 hours 7.7 Can support handswitch and remote control exposure 8. The machine parameters 8.1 Display size: 8 inch display capacitive touch screen 8.2 Whole machine width: 240 mm 8.3 Whole machine length: 440 mm 8.4		
			Whole machine height: 236 mm 8.5 Whole weight(no detector): 18kg 9. Options 9.1 Wireless network data transmission (WIFI) 9.2 Choose a chest rack 9.3 The selection of moving flat bed 10. After sale service 10.1 The machine warranty 1 year 10.2		
			Provide complete operation manual 1 set 10.3 Free service hot line 11, copyright certificate 11.1 Utility model: a kind of medical product damping rotation axis 11.2 Appearance patent: portable digital X-ray machine 11.3 Software copyright: Browiner technology portable X-ray machine embedded system software		
15	6	unit	Wheeled Stretcher - STANDARD	58,200.00	349,200.00
			Product size: 190cmx54x88cm, Size of castors: 160cm, N.W.:35kg, Load bearing<<< 129kg		
16	1	unit	Infant Incubator	624,195.00	624,195.00
			Power supply: AC220V-230V/50Hz or AC110-120V/50-60Hz; Max Power Output: 850VA (SK-N004) 1000VA (SK-N005); Mode of Control: Air mode and baby mode controlled by micro-computer; Air temperature control range: 25°C - 37°C (override mode 37-39°C); Baby temperature control range: 34 - 37°C (override mode 37-38°C); Temperature variability: ≤0.5°C; Temperature uniformity: ≤0.8°C; Noise inside hood: ≤45dB(A) (Environment noise ≤35dB(A)); Over-temperature alarm: Yes; Sensor alarm: Yes; Power failure alarm: Yes; Sensor box position alarm: Yes; System alarm: Yes; Accuracy of humidity: ±10%RH; Humidity display range: 0%RH-99%RH; Humidity control range: 0%RH-99%RH		
17	1	unit	LED O.R. Light	485,000.00	485,000.00
			Light Intensity at 1m distance: 160,000lx		
			Dimmability: 32,000-160,000lx		
			Colour temperature: 4,700K		
			General Colour rendering index: R 98		
			Colour rendering index R (red): 99		
			Light field diameter: D10=215mm, D50=120mm		
			Adjustable light field diameter: 215-290mm		
			Depth of illumination: 60% 450mm, 20% 750mm		
			Temp. of light head surface:<27.5°C		
			Total irradiance at 100,000lx: 325W/m ₂		
			Power input of light head (100%light): 56watt		
			Endoscopy mode: Green light		
			Service life of lamp: >40,000h		
			Degree of protection of light head: IP 54		
			Light: cool light with no infrared components; optimum illumination in area and depth		
			Shadow-free with the ideal ratio of lumen to light-field diameter,		
			Natural color rendering index with white light (Ra of 98 at all light levels, R9 of 99 in the red region)		
			Can be regulated from 32,000 lux to 160,000 lux		
			Hygiene:		
			<i>Page 6 of 7 pages ...</i>		
			<i>Next page pls ...</i>		



Republic of the Philippines
 PROVINCE OF BOHOL
 City of Tagbilaran



OFFICE OF THE GOVERNOR

NOTICE OF AWARD

November 5, 2020

F-NA-2020-851
 11/06/20

MS. STELLA MARIE C. ANCLA
 Proprietor/Manager
DEXTEL TRADING
 0410 Tamblot Street, Cogon District,
 Tagbilaran City

Dear Ms. Ancla,

We are pleased to inform you that the Procurement of **MEDICAL EQUIPMENT FOR HOSPITAL MODERNIZATION** of the **PROVINCIAL HEALTH OFFICE, NEW CAPITOL BUILDING, COGON DISTRICT, TAGBILARAN CITY** under **PROJECT REFERENCE NO. – PHO – 2020 - 039** is hereby awarded to you as the Bidder with the Single Calculated and Responsive Bid with a Contract Price equivalent to **TWENTY THREE MILLION FIVE HUNDRED THIRTY SIX THOUSAND FIVE HUNDRED SIXTY FIVE PESOS ONLY (P 23,536,565.00).**

In view thereof, you are hereby issued this Notice of Award subject to the submission to this Office within ten (10) days from receipt hereof, a Performance Security in any of the following forms:

a.	Cash, Cashier's Check Manager's Check, Bank Draft or Irrevocable Letter of Credit (5% of the total amount of the Purchase Order)	Php 1,176,828.25
b.	Bank Guarantee (5% of the total amount of the Purchase Order)	Php 1,176,828.25
c.	Surety Bond (30% of the total amount of the Purchase Order)	Php 7,060,969.50

Failure on your part to submit the required Performance Security within the period specified shall be sufficient cause for the cancellation of the Award and forfeiture of your Bid Security.

Within this notification, you are required to prepare the items awarded to your company for delivery as indicated in the Purchase Order. Failure to deliver within the stipulated time will mean imposition of liquidated damages and other sanctions that may deem fit pursuant to the provisions of Republic Act 9184.

Very truly yours,

ATTY. ARTHUR C. YAP
 Governor

By Authority of the Governor:

K. Pioquinto
ATTY. KATHYRIN FE D. PIOQUINTO
 Provincial Administrator

Date Approved: **09 NOV 2020**

I acknowledge receipt of this Notice:

Date Released to Supplier: **NOV 11 2020**

Date Returned by Supplier: **NOV 11 2020**

: _____
 Authorized Signature
 (Name of the Representative of the Bidder)

: _____
 Authorized Signature
 (Name of the Representative of the Bidder)

PHO-255

771

OK



BAC Resolution Declaring Single Complying and Responsive Bid and Recommending Approval

RESOLUTION NO. 2020-289
Series of 2020

WHEREAS, the **PROVINCIAL GOVERNMENT OF BOHOL** caused the advertisement and posting of Invitation to Bid in the Newspaper, PHILGEPS Website, Provincial Government Website and at three (3) conspicuous places in the premises of the Provincial Capitol Complex, Tagbilaran City of the hereunder project:

Name of Project : **PROCUREMENT OF MEDICAL EQUIPMENT FOR USE OF THE HOSPITAL MODERNIZATION**
Project Reference No. : **PHO – 2020-039**
PMO/End User : **PROVINCIAL HEALTH OFFICE, NEW CAPITOL BUILDING, TAGBILARAN CITY**
PMO/ABC : **P 24,264,500.00**

WHEREAS, during the opening of bids on October 14, 2020, the following suppliers **DEXTEL TRADING** drop its bid proposal for the said procurement;

WHEREAS, after the opening of bids, the following are the result of the bid proposals of the bidders to wit:

Name of Bidder	Approved Budget for the Contract (ABC)	Bid amount	Variance from the ABC
DEXTEL TRADING	P 24,264,500.00	P 23,536,565.00	P 727,935.00

WHEREAS, in the detailed evaluation of bids **DEXTEL TRADING** was declared as the Single Calculated Bid as Read in the amount of **P 23,536,565.00**;

WHEREAS, upon the conduct of post qualification by the Technical Working Group, **DEXTEL TRADING**, met the required legal, technical and financial requirements for the said project;

WHEREAS, upon recommendation by the Technical Working Group and verification, validation and ascertaining of all statements made and documents submitted by **DEXTEL TRADING** the bidder declared as the **SINGLE CALCULATED BID**, successfully passed the Post Qualification process, thus the Committee now considered its bid as the **SINGLE CALCULATED and RESPONSIVE BID** for the said project;

NOW THEREFORE, We the Members of the Bids and Awards Committee, hereby RESOLVE, as it is hereby RESOLVED:

- a.) To declare **DEXTEL TRADING** as the **SINGLE CALCULATED and RESPONSIVE BID** in the amount of **P23,536,565.00** for the Procurement of **MEDICAL EQUIPMENT FOR USE OF THE HOSPITAL MODERNIZATION** under Project Reference No.- **PHO - 2020 – 039**;
- b.) To recommend award to **DEXTEL TRADING** in the contract amount of **TWENTY THREE MILLION FIVE HUNDRED THIRTY SIX THOUSAND FIVE HUNDRED SIXTY FIVE PESOS ONLY (P 23,536,565.00)**;
- c.) To recommend for the approval by the **HONORABLE GOVERNOR** of the **PROVINCIAL GOVERNMENT OF BOHOL** the foregoing findings.

RESOLVED, at the OFFICE THE PROVINCIAL PROCUREMENT MANAGEMENT UNIT, CAPITOL COMPLEX, TAGBILARAN CITY, this 28th day of October 2020.



NILO G. AHAT
Provincial Legal Officer
BAC-Chairman




GISELLE F. QUIMPO
Chief of Staff - GO
BAC Vice-Chairman



PETER ROSS M. RETUAL
Provincial Budget Officer
BAC Member



LARRY M. PAMUGAS
OIC – Provincial Agriculturist
BAC Member



CESAR TOMAS M. LOPEZ
Provincial Health Officer I
BAC Member




CAMILO L. GASATAN
Provincial Engineer
BAC Member



CESAR TOMAS M. LOPEZ
PMO/End User

APPROVED: **BY AUTHORITY OF THE GOVERNOR**


ATTY. KATO YON FE D. PIOQUINTO
PROVINCIAL ADMINISTRATOR

ATTY. ARTHUR C. YAP
Governor
Approved on _____ 