

Revised on: May 24, 2004

Standard Form Title: Purchase Order

PURCHASE ORDERPROVINCIAL HEALTH OFFICE, NEW CAPITOL BUILDING, TAGBILARAN CITY

Agency/Procuring Entity

18036-14

Supplier	DEXTEL TRADING	D.O. No. :	P.O. # PHO - 2021 - 336
Address	0410, Tamblo Street, Cogon District, Tagbilaran City	Date :	December 13, 2021
Telephone No.:		Mode of Procurement :	Public Bidding
TIN			

Gentlemen/Mesdames:

Please furnish this Office the following articles subject to the terms and conditions contained herein:

Place of Deliver	PROVINCIAL HEALTH OFFICE, NEW CAPITOL BUILDING, TAGBILARAN CITY	Delivery Term: Please read at the back hereof
Date of Delivery	Twenty five (25) days after receipt of NTP	Payment Term: Upon Completion of Delivery
Warranty Period	ONE(1) YEAR	

Item No	QTY	UNIT	DESCRIPTION OF SUPPLIES	UNIT COST	AMOUNT
1	5	unit	X-Ray Digitalization, GOS, Preview: <2sec., Cycle: 6-9sec, 16 bits per pixel, 150um, pixel size: 2,836x2,336 / 2,836x2,832, Image Area Size (in.): 16.8x13.8 / 16.6x16.6", Wired use possible with SE detector cable & MP box Wireless: FP D to console, in room tyup. 33ft range, closed loop, image data only (no patient info). IEEE 802.11in, 2.4 & 5.2 GHz bands, W52/W53/W56/W58) WPA2-PSK encryption w/AES & MAC (unique IP) protocols secure connection, - CHINA/BROWNER/FUJI	3,475,000.00	17,375,000.00
			confirmatin & completion of data, handshake pairing to registered FPDs only; LAN wire Ethernet: 10/100/100 Base-T DHCP or Static or Wireless IEEE 802.11 (depending on HOsp. network); Battery: User swappable Lithium ION up to 4 hrs or 1400 images (200 continuous images). standby: 4 hrs, sleep: 7.5, deep sleep: 18.5hrs; 40 lines/cm recommended: 40-44 lines/cm, 80 or more lines/cm useable processing (option) simulates scatter clean up for images acquired without a grid, 352lbs point load; 683lbs distributed; (protective cover required for standing exams; IPX-6, Internatinal protection classification, (not a substitute for normal precautions such as bagging.)		
			Resistance effectiveness can be impaired by wear and drops. FPD cable connector must be fully covered with adapter.; All: single phase 50/60Hz Std 110VAC FPD MP box: 100-240VAC (+/-10%), 200VA or less, console: 115/230VAC (auto), 460VA or less Battery charger: 100-240VAC (+/-10%), 100VA or less; Heat Output: FPD:316KJ/h console: 324KJ/h (standby); includes detector(s), Battery(s), FPD cable, MP box, Handswitch Interface & FDX console technologist workstation.		
2	1	unit	Digital X-Ray, robust and maneuverable with wide range of movements; 1 unit of wireless digital panel detector in between both wall stand and exam table; flat panel detector provided must be form gadolinium oxysulfide scintillator and/or amorphous silicon type photo-detector; single central/main console unit; - BROWNER SDX 5000	3,450,000.00	3,450,000.00
			color display preview of images (≤6 seconds); easy to operate, patient friendliness, high quality images; specified country of origin with documents; multi-pulse high frequency inverter at least 120kHz; power output (at least 50kW; 500mA at 80kV in 0.1s) mA range selection: 10 to 500 mA. X-ray Tube: high speed rotating anode typ X-ray tube; tube voltage operating range: 40 to 125kV;		
			Dual focus spot size (small focus ≥ 0.6mm; large focues ≤ 1.2mm); X-ray Tube support and collimator: floor mounted free-standing system, lage color LCD touch panel on the tube support unit with set/change of the following: technique, s-ray tube voltage, s-ray tube current, automatic exposure, x-ray tube rotation rage: angulation: -135° to +135° loct at 0° &±90 °; all trainings shall be conducted with two weeks from acceptance; 1 year warranty; certification from the suppliers that the spare parts are available for at least 5 years fromthe date of the warranty period.		