

PURCHASE REQUEST  
Province of Bohol



Health PR. No. 2022-102  
SAI No.  
OBR No.

Date: March 10, 2022  
Date:  
Date:

Item No.	Quantity	Unit of Issue	SAI No.	Estimated Unit Cost	Estimated Cost
6	1	unit	<p><b>V. OPERATING SYSTEM AND STORAGE</b> 1. Linux Operating System with Hard Drive memory of not less than 500 GB and up to 1TB.</p> <p><b>VI. TRANSDUCERS</b> The probes shall include multi frequency imaging capability for optimal resolution and penetration. The endocavity probe should have temperature detection technology with indication of real time temperature of the probe head on the monitor 1. Convex Probe (1-7 MHz). Clinical Applications: Abdominal Obstetrics and Gynecology. 2. Endocavity Probe (3 to 15MHz). Clinical Applications: Gynecology, Obstetrics, Urology. With no less than 190-degree FOV and temperature detection technology, can support elastography function for endocavity checking. 3. Linear Probe (4-16MHz) Clinical Applications: Vascular, Small parts, Musculoskeletal, Neurology</p> <p><b>VII. PATIENT DATA MANAGEMENT</b> 1. The system must have built-in patient and report review tool 2. The system must support JPG, BMP and TIF Image format 3. The system must support AVI and WMV Video format 4. The system must support PDF Document format 5. Must have Audio output and input, Network port, MIC input, Video input and S-video input. 6. Must have 6 USB port and 1 engineering port</p> <p><b>VIII. ACCESSORIES</b> 1. B/W Thermal Printer 2. 2 KVA Online, Double Conversion UPS</p> <p><b>PATIENT MONITOR (Cardiac Monitor)</b></p>	350,000.00	350,000.00