

PURCHASE REQUEST
Province of Bohol



Health Ref. No: 2022-102
SAI No.
OBR No.

Date: March 10, 2022
Date:
Date:

Quantity	Unit of Issue	SAI No.	Estimated Unit Cost	Estimated Cost
7	1 unit	<p>4. SpO2</p> <ul style="list-style-type: none"> i. Range: 0 ~ 100 % ii. Accuracy: <ul style="list-style-type: none"> ☑ Adult/Pediatric: ±2% (70%~100% SpO2) ☑ Neonate: ±3% (70%~100% SpO2) iii. Pitch Tone iv. PR range: 25 bpm ~ 300 bpm v. PR Accuracy: ±2 bpm vi. PI: 0-10 <p>5. TEMP</p> <ul style="list-style-type: none"> i. Position: Skin, oral cavity, rectum ii. Channel: 2 iii. Measuring Range: 0°C ~ 50°C iv. Accuracy: ±0.3°C v. Transient Response Time: ≤30 s <p>III. POWER SUPPLY</p> <ul style="list-style-type: none"> 1. The supply must be 100~240 V AC, 50/60 Hz 2. The battery should be Li-ion rechargeable with 5000 mAh that can long up to 6 hours or better <p>IV. ACCESSORIES:</p> <ul style="list-style-type: none"> 1. One (1) pc reusable ECG cable 2. One (1) pc adult reusable SpO2 sensor 3. One (1) pc adult reusable NIBP cuff 4. One (1) pc reusable NIBP tube 5. One (1) pc reusable skin temperature probe <p>SUCTION MACHINE, HEAVY DUTY</p> <p>I. FEATURES</p> <ul style="list-style-type: none"> 1. The machine can be widely applied in the surgical operations which need to absorb phlegm, thick liquid, etc. <p>II. TECHNICAL SPECIFICATIONS</p> <ul style="list-style-type: none"> 1. The power requirement should be 220 V, 50/60 Hz 2. The operation noise of the pump should be less than or equal to 65 dB(A) 3. It has 40 L/min pumping frequency 4. Maximum negative pressure should be greater than 90 kPa 5. It has an input power of 180 VA <p>III. ACCESSORIES</p> <ul style="list-style-type: none"> 1. Two (2) pcs Reservoir, 2,500 mL Capacity x x x <p>charged to MDFO Loan</p> <p>Terms of reference:</p> <ul style="list-style-type: none"> 1. Bidder must provide certificate of exclusive or authorized distributorship for the equipment being offered and that they are the authorized distributor for this bid. 2. Bidder must have track record of having supplied and successfully installed various medical equipment in the following: a) DOH-retained hospital, b) two (2) LGU hospitals and private institutions. 	250,000.00	250,000.00