Total to of Component LOUs with the confidence of Component LOUs start in the componen

magement of /

PURCHASE REQUEST Province of Bohol



Equipment for the use of hospital in excess of the ABC) shall be

vent: Provincial Health Office Section:			PR. No: 2022-102 SAI No. OBR No.	Date: Date: Date:	March 10, 2022
Item No.	Quantity	Unit of Issue	SAI No.	Estimated Unit Cost	Estimated Cost
1	1	unit >	OPERATING ROOM LIGHT, DUAL-HEAD Features: 1. Reflection system should be made of CAD/CAM technology or better 2. Illumination depth should be up to 1,300 mm or greater 3. Panel should made of PC plastic or better 4. Should be 360° universal design that can be satisfy various needed	2,300,000.00	2,300,000.00
F-360			height and angle in the operation. II. TECHNICAL SPECIFICATIONS 1. Illuminance should be at least 160,000 lux and 120,000 for each lamp head. 2. Color temperature should be 3,800 K, 4,400 K and 5,000 K 3. 95% Color Reduction Index or better 4. The total irradiance should be 450 W/m2 or better 5. Light field size should be adjustable from 160 mm up to 280 mm or better 6. The light source should be LED and the service life of the light source should be 50,000 hours or greater		
Wiest gongoes /			7. The power consumption of the equipment should be 60 W or lower for each arm 8. Total LED bulb quantity should be at least 180 pcs or greater 9. Should be 8-stage continuous bright light adjustment or better 10. The installation height of the operating light should be at least 2,900 mm or better III. ACCESSORIES 1. One (1) pc Sterilizable Handle		WAY CENEMA
2	1	unit	CAUTERY MACHINE (ELECTROSURGICAL UNIT)	1,200,000.00	1,200,000.0