

PURCHASE REQUEST  
Province of Bohol



viewed as to availability of funds  
PETER ROSS M. RETUTAN  
Provincial Budget Officer

PR. No: 2022-102  
SAI No.  
OBR No.  
Date: March 10, 2022  
Date:  
Date:

| Item No. | Quantity | Unit of Issue | SAI No.   | Estimated Unit Cost | Estimated Cost |
|----------|----------|---------------|---|---------------------|----------------|
|          |          |               | <p>ELECTROSURGICAL UNIT<br/>Brand: LED<br/>Model: Surtron 300HP</p> <p>1. FEATURES</p> <p>1. Twin operative activity</p> <p>i. Two output channels for monopolar cutting</p> <p>ii. Two independent setting and output channels for the most frequent monopolar coagulation</p> <p>2. Controlled cutting process</p> <p>i. Enhanced cutting current for eschar free incision</p> <p>ii. Cutting current with selectable coagulating property for bleeding-controlled dissection</p> <p>iii. Low capacitive coupling by slow BLEND current for endoscopic procedures</p> <p>3. Twin controlled coagulative procedures</p> <p>i. Speedy/Fulgurate/Forced or equivalent coagulation function and Deep/Pin Point/Desiccate/Soft or equivalent coagulation function are available to two operators</p> <p>ii. Spray or equivalent coagulation function available by hand switch or foot switch controls</p> <p>4. Wide bipolar features</p> <p>i. Bipolar cutting current with selectable bleeding control</p> <p>ii. Slow BLEND cutting current</p> <p>iii. Separately activatable Automatic Start/Stop coagulation</p> <p>5. Surgeon and patient total safety</p> <p>i. Acceptability of contact impedance between patient body and return electrode continuously watched by Skin Plate Electronic Control (SPEC) or equivalent function</p> <p>ii. High frequency (HF) leakage currents continuously monitored</p> <p>iii. Mains fluctuation automatically compensated before to start the delivery of HF power</p> |                     |                |



PROVINCIAL HEALTH OFFICE