

**PURCHASE REQUEST**  
Province of Bohol



Provincial Health Office

PR. No: 2022-102

SAI No.

OBR No.

Date: March 10, 2022

Date:

Date:

Item No.	Quantity	Unit of Issue	SAI No.	Estimated Unit Cost	Estimated Cost
4	1	unit	<p><b>FEATURES:</b></p> <ul style="list-style-type: none"> <li>• Mechanical structure for height adjustment, steel crank</li> <li>• One button lay down PP side rail, gas spring controlled</li> <li>• PP surface easy to clean, with removable mattress</li> <li>• Backrest raising with CPR release, gas spring controlled</li> <li>• Central brake system with both sides control, 150 mm diameter</li> <li>• Steering wheel technology, shock-proof</li> <li>• With I.V. pole socket at both sides of head part and oxygen tank holder</li> <li>• Storage tray at the base</li> </ul> <p><b>TECHNICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>• 250 kg maximum capacity</li> <li>• Dimension: 1930 mm x 640 mm x 500-900 mm</li> <li>• Backrest raising: 0-85°</li> <li>• Height adjustment: 400 mm</li> </ul> <p><b>FETAL AND MATERNAL MONITOR (Cardiotocogram)</b></p> <p><b>PHYSICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>• Power Requirement: 100~240 V AC; 50/60 Hz</li> <li>• 12.1" LCD touch screen display; 800 x 600 pixels resolution</li> <li>• RS232 interface and RJ45 interface</li> <li>• Ultrasound Transducer                             <ul style="list-style-type: none"> <li>o Cable Length: 2.5 m</li> <li>o Dimension: 88 mm x 35 mm</li> </ul> </li> <li>• TOCO Transducer                             <ul style="list-style-type: none"> <li>o Cable Length: 2.5 m</li> <li>o Dimension: 88 mm x 35 mm</li> </ul> </li> <li>• Remote Event Marker Cable Length: 2.5 m</li> <li>• ECG Cable Length: 3 m</li> <li>• SpO2 Cable Length: 2 m</li> <li>• NIBP Cable Length: 3 m</li> <li>• Temperature Probe Cable Length: 3 m</li> <li>• Battery Working Time: ≥2 hours</li> <li>• Built-In Printer Recorder</li> </ul> <p><b>PERFORMANCE SPECIFICATIONS</b></p> <ul style="list-style-type: none"> <li>• FHR Parameter                             <ul style="list-style-type: none"> <li>o Fetal Heart Rate Measurement: 50~ 240 bpm; 1 bpm resolution; ±2 bpm accuracy</li> </ul> </li> <li>• Ultrasound Output                             <ul style="list-style-type: none"> <li>☑ Isppa.3&lt;190 W/cm2</li> <li>☑ Ispta.3&lt;94mW/cm2</li> <li>☑ Isata&lt;20 mW/cm2</li> <li>☑ TI&lt;1.0</li> <li>☑ MI&lt;1.0</li> </ul> </li> </ul>	650,000.00	650,000.00