

REPORT

PURCHASE REQUEST
Province of Bohol



PTR No.:
 Date: May 25, 2022
 EQUIP0012022

Provincial Health Office: _____
 PR. No.: 2022-102
 SAI No.: _____
 OBR No.: _____
 Date: March 10, 2022
 Date: _____
 Date: _____

Item No.	Quantity	Unit of Issue	SAI No.	Estimated Unit Cost	Estimated Cost
5	1	unit	<ul style="list-style-type: none"> o SpO2 Alarm Limits: 50%~100% • NIBP Parameter o Measurement: Systolic Pressure, Diastolic Pressure, Mean Artery Pressure o Method: Oscillometric o Measurement Range: <ul style="list-style-type: none"> <input checked="" type="checkbox"/> Systolic Pressure: 40~235 mmHg <input checked="" type="checkbox"/> Diastolic Pressure: 10~215 mmHg <input checked="" type="checkbox"/> Mean Artery Pressure: 20~235 mmHg o Measuring Time: ~30 seconds to 45 seconds o Cuff Pressure Measuring Range 0~300 mmHg • TEMP Parameter o Measurement Range: 0°C~50°C; 0.1°C resolution o Temperature Unit: °C, °F o The Refresh Time 1~2 s o Self-Check Test: 5~10 min o Measurement Mode: Direct Mode <p>ACCESSORIES</p> <ul style="list-style-type: none"> • Two (2) pcs US Transducer • One (1) pc TOCO Probe • One (1) pc Marker • Three (3) pcs Transducer Belt • One (1) pc ECG Cable • One (1) pc SpO2 Sensor • One (1) pc NIBP Cuff • One (1) pc NIBP Tube • One (1) pc Temperature Probe <p>CONSUMABLES</p> <ul style="list-style-type: none"> • One (1) pc Ultrasound Gel • Fifty (50) pcs ECG Electrodes, Disposable • Ten (10) pcs Recording Paper <p>ULTRASOUND MACHINE, CONSOLE, GENERAL IMAGING</p>	3,400,000.00	3,400,000.00

RECEIVED