

PURCHASE REQUEST
Province of Bohol



PACCO	Documentation	Cash availability, prepare	Approval of DV, signing of check	Returned to Dept./Office concerned
PTO				
GO				

Office: **Medical Health**
PR. No: **2022-102**
SAI No. _____
OBR No. _____

Date: **March 10, 2022**
Date: _____
Date: _____

Item No.	Quantity	Unit of Issue	SAI No.	Estimated Unit Cost	Estimated Cost
			<p>I. SYSTEM DESIGN</p> <ol style="list-style-type: none"> 1. The Ultrasound system have B, Simultaneous Dual, M, Color M, Anatomical M-mode, Doppler (Color, Spectral and Power) and triplex modes. 2. Must have 4 active ports, 1 parking port and at 6 probe holders. 3. Have at least 21.5" LED articulating Monitor at minimum resolution of 1920 x 1080 pixels with Anti-Flickering technology and manual adjustment for Contrast and Brightness. 4. Must have Angle adjustable of 20 degrees with control panel adjustable Up and Down 0-230 mm, swivel angle of ±40 degrees. 5. Must have at least 13.3" User Configurable Touch Screen Control Panel and Digital QWERTY Keyboard. 6. Must have built-in/integrated CD-R/DVD-R and at least 7 USB ports including the Engineering port. 7. Must have full DICOM capability with LAN connectivity and WLAN capability. 8. Must have built-in battery with at least contentions working time of 90 minutes. 9. Must have built-in Gel Warmer. 10. Must have Rear, Front Handles for easy transport. 11. Must have a unit size of Depth: 700-800 mm; Width: 500-530 mm; Height: 1110-1680 mm 12. Must have a weight not more than 80 kg 13. Must have a boot up time of ≤60 s and shutdown time of ≤20 s. 		