

GMPH PR # 3-22-162

Date: March 25, 2022

Department: GMPH

PR No. _____

Date: _____

Talibon, Bohol

SAI No. _____

Date: _____

Section: _____

ObR No. _____

Date: _____

Item No.	Quantity	Unit of Issue	Item Description	Estimated Unit Cost	Estimated Total Cost
1	2	units	Cardiotopography (CTG) Machine with UPS/ Ultrasound Fetal Monitor with UPS Compact design, it can be placed horizontally or hung on the wall; color LCD, 60° convertible screen; Display monitoring curve and data clearly; Area mark for normal FHR (120 BPM~160BPM); Manual record fetal movement; Alarm for abnormal FHR; Real-time monitoring for continuous 24 hours; Storage, playback and print the monitoring curve for continuous 12 hrs; Optional interface languages: Chinese and English; Optional twins monitoring; 9-crystal and board beam transducer; Extra long-life high-resolution built in thermal printer; Communicate with central monitoring system by built in RJ45; Type of protection against electric shock: Class I; Degree of protection against electric shock: type B applied part; Working voltage: AC100V ~ 240V; Frequency: 50Hz/ 60Hz P<60VA; Fuse: T1.6AL250V; Record print paper Z-fold paper; print width 112mm; Valid print width: 104mm; Paper speed: 1cm/min, 2cm/min, 3cm/ min (optional) Ultrasound probe nominal frequency: 1.0MHz; Work Frequency: 1.0 MHz +/-10%; Negative peak sound pressure: p<1 Mpa; Output beam intensity: lob<20 mW/cm2; The peak time space peak intensity: Ispta<100mW/cm2; The average time space peak intensity: Isata <10 mW/cm2; FHR Range: 50BPM~240BPM; Resolution : 1BPM; Accuracy: +/-2 BPM; TOCO Range: 0~100%; Resolution: 1%; Non-linear error <- 10%; RZ way: manual; Fetal marking: Manual button; Package Included: 1 x CMS800G Fetal Monitor, 1 x 3 in 1 Transducers (Ultrasound Transducer, TOCO Transducer, Remote Marker); 2X Abdomen belts; 1x Earth line; 1 x Power cord; 2x Fuses; 10 sets X Record Paper; 1x User manual	110,000.00	220,000.00
Page 1 of 4					
				SUB-TOTAL....	220,000.00

Purpose: For GMPH Various Hospital Ward Use

Condition: 1. To be delivered at GMPH, Talibon, Bohol

4. The supplier discloses the brand and/or manufacturer.

2 To be delivered 15 days after approval of P.O.

3 Public Bidding

BY AUTHORITY OF THE GOVERNOR

Requested by:	Cash Availability	Approved by:
Signature Printed Name Designation MIGUELITO S. JAYOMA, MD.,FPSGS,FPSCS,MHM. Chief of Hospital	MR. EUSTACIO A. SOCORIN Provincial Treasurer	ATTY. KATHYRIN FED. PIOQUINTO PROVINCIAL ADMINISTRATOR HON. ARTHUR C. YAP Provincial Governor

Charge to 20% DF 2022
Medical EquipmentGMPH-HOPSS-PSS FO - 01
REV. 0 October 01, 2017