

Flow-optimised luminaries housing for laminar airflow ceiling  
 Degree of protection IP 54 as per IEC 144  
 Suitable for Spray and wipe disinfection  
 See Better:  
 Endoscopy mode  
 Improves the contrast for the surgeon using imaging procedures  
 Reduced reflection on monitors during endoscopic procedures  
 Provides adequate background lighting for orientation for the complete surgical team  
 Maintenance & Energy:  
 Technology for 100% lighting power even after many years of use

**CONTROL ELEMENTS:**

Simple and clear structure of control elements  
 All functions are quickly and accurately accessible  
 Disinfectant resistant surfaces (easy to clean)  
 Adjustable illuminance  
 Endoscopy mode  
 Head 1 ( main Light) + Head 2 (Satellite)  
 Maximum illuminance 160,000 lux with constant colour temperature 4,700K  
 Colour Rendering index CRI (Ra) = 98 at all dimming -steps  
 Colour Rendering index for red colours R9=99 for natural and detailed rendition of tissue structure  
 Switch-over for endoscopic illumination (green light)  
 Dimmable in 8 levels  
 Diameter of shadow less light field adjusted via centrally

x-x  
**TOTAL** > > >


23,536,565.00  
 WWWWWW

Page 7 of 7 pages...

(total amount in words) **TWENTY THREE MILLION FIVE HUNDRED THIRTY SIX THOUSAND FIVE HUNDRED SIXTY FIVE PESOS ONLY**

In case of failure to make the full delivery within the time specified above, a penalty of one-tenth (1/10) of one (1) percent for everyday of delay shall be imposed.  
 Conforme:

**DEXTEL TRADING**  
 Signature over printed name of Supplier  
NOV 25 2020  
 Date

Very truly yours,  
  
**ATTY. ARTHUR C. YAP**  
 Governor

**ATTY. KATHYRIN FE D. PIOQUINTO**  
 Provincial Administrator  
 Date approved: \_\_\_\_\_

Funds Available: \_\_\_\_\_  
**EUSTAQUIO A. SOCORIN**  
 Provincial Treasurer

Earmarked No. : \_\_\_\_\_  
 Amount : 2197  
P 24,264,500.00